
* Hanger Clinic

Sock Fitting Guide


It is important to re-evaluate the fit of your socket every morning, afternoon, and evening.



NOT ENOUGH SOCKS

- Leg will fall into socket causing pressure on bottom of leg and base of knee cap.
- · Prosthetic leg may rotate.


<u>Pin-lock:</u> All clicks will engage fast and when seated.



TOO MANY SOCKS

- · Knee will not be seated fully in socket.
- May experience "hammocking" pressure on bottom of leg due to pressure of socks pulled tight.

<u>Pin-lock:</u> You may not hear any clicks when seated or clicks are a struggle to achieve when seated and standing.



JUST RIGHT

 Global pressure with no specific areas of tightness or pain.

<u>Pin-lock:</u> You will hear a few clicks when seated but most clicks when you stand and take first few steps.



